

**Govt of West Bengal**  
**Office of the Chief Medical Officer of Health**  
**Cooch Behar**

Tel: 228874 (03582) Fax: 228966

E-mail: [cmoh\\_cbr@wbhealth.gov.in](mailto:cmoh_cbr@wbhealth.gov.in)

**NIT NO-CMOH/CBR/2020-2021/HWC/03**

No: - 3828

Date: - 27/07/2021

**E.TENDER NOTICE**

E.Tenders are hereby invited from the reputed suppliers having valid trade license for supply Laboratory Items. Intending bidders are requested to download the tender document from the website [www.wbhealth.gov.in](http://www.wbhealth.gov.in), [www.wbtenders.gov.in](http://www.wbtenders.gov.in) & [www.coochbehar.gov.in](http://www.coochbehar.gov.in) in on and from 28.07.2021 and indicate clearly the price both in figure and words inclusive of all taxes. Quotated rate will be valid upto 31.03.2022.

An amount of Rs.18,660 (Eighteen thousand Six hundred Sixty) is to be given through Online payment. Deposition of earnest money must be through a) Net Banking (any of the banks listed in the ICICI Bank gateway) in case of payment through ICICI bank payment gateway. b) RTGS/NEFT in case of offline payment through bank account in any bank. Supply order will be placed as and when required basis through the period.

Price quoted should be for units specified and delivery at the offices mentioned at the time of order (sub centre wise packaging will be required). All risks to goods till the point of delivery shall be on the supplier. The quoted price should be inclusive of all charges GST, freight, packing forwarding and insurance. In case of delay supply items 5% will be received as liquidated damage. Details of items are attached as Annex-A

Sl no	Particulars	Date & Time
1	Date of publishing N.I.T. & other Documents online	: 28.07.2021, 10 AM
2	Online documents download start date	: 28.07.2021, 10 AM
3	Online documents download end date	: 10.08.2021, 5PM
4	Online Bid submission start date	: 28.07.2021, 10 AM
5	Pre-bid meeting to be held at Office of the Tender Inviting authority	: 04.08.2021, 2 PM
6	Dates for submission of Earnest Money Deposit online	: from 28.07.2021, 10 AM, to 10.08.2021, 5PM
7	Online Bid Submission closing date and time	: 10.08.2021, 5PM
8	Online Bid opening date for Technical Proposals	: 13.08.2021, 10 AM
9	Date of online uploading list for Technically Qualified Bidders	To be notified later
10	Date of online opening of Financial Proposal	To be notified later
11	Sample verification	To be notified later

Document Details mentioned below:-

SL. No.	Category Name	Sub Category Description	Details
1	Certificate(S)	Certificate(S)	Income Tax PAN, Professional Tax registration/ any Challan deposit in last six months from schedule date of tender opening. GST Registration along with copy of last return filed.

Company Details (s)	Company Details	Certificate of incorporation/Partnership Deed. Updated Trade License, Drug license (where applicable) Power of Attorney in favour of signatory of bid.
3	Credential - 1	Performance statement. Credential for last 3 years.
4	Declaration - 1	Audited Balance Sheet & Profit & Loss A/c of last 1 years. Name, address of banker, account no.

  
CMOH & Secretary,

Dist. Health & Family Welfare samiti, Cooch Behar

Ho  
27/7/24

Dated, Cooch Behar, the

Memo No: -

Copy forwarded for information and wide publication at his office notice board please to:-

1. The Director of Health Services, West Bengal, Swasthya Bhawan, Kolkata-700091.
2. The Sabhadhipati, Zilla Parisad, Cooch Behar.
3. The District Magistrate & Ex-Officio, Executive Vice Chairman, D.H. & F.W. Samiti, Cooch Behar.
4. The District Information & Cultural Affairs Officer, Cooch Behar.
5. The Chairman, Cooch Behar Municipality.
6. The Post Master, Cooch Behar Head Post Office, Cooch Behar.
7. The Station Master, Cooch Behar Rly. Station, Cooch Behar.
8. The Dy. Chief Medical Officer of Health- II Cooch Behar.
9. The Dy. Chief Medical Officer of Health- III Cooch Behar.
10. The Dist. Leprosy Officer, Cooch Behar.
11. The Dist. Tuberculosis Officer, Cooch Behar.
12. The Superintendent, M.J.N. Hospital, Cooch Behar.
13. The Accounts Officer of this office.
14. The Officer NOTICE BOARD of this office.

  
CMOH & Secretary, Dist.

Health & Family Welfare Samiti, Cooch Behar

Ho  
27/7/24

# Equipments specification for Suswastya Kendras (Non Cat Items)

Annex - A

SL. NO.	ITEMS	SPECIFICATION OF ITEMS	Total Number of Items to be PROCURED
1.	CUSCO'S/GRAVES SPECULUM VAGINAL BI VALVE MEDIUM	Cusco's / graves Speculum Vaginal bi - Valve Medium:- a) Material: Stainless Steel b) Colour: Silver c) Usage/Application: Clinical d) Size : (Medium) 85 x 30 mm e) Precisely designed f) Optimum finish g) Corrosion proof	73 No
2.	FOREIGN BODY HOOK (ENT)	Foreign Body hook (ENT): a) Categories: Holding Instruments b) Type: Non - Disposable c) Reusable: Yeas d) Color: Silver e) Material: Stainless Steel f) Usage: Multiple g) Condition : New	73 No
3.	DIGITAL FOETAL DOPPLER	FOETAL DOPPLER a) It should be LCD Display (45mm X 22mm or more) b) It should be ultrasonic emitting frequency: 2 - 2.5 MHz and power: < 5m Watt /cm2 c) Overall sensitivity at the distances 120 - 200 mm from the face of the probe: 90db d) Automatic FHR test range 50 - 210 BPM or wider (Beat / Minute) e) Resolution: 1 BPM f) Audio output : 0.5 - 1 Watt g) Should be supplied with electric charger & rechargeable battery (Li-ion) with 1 set of spare battery. h) Good quality water proof case for carrying purpose i) Good quality water proof probe & gel (5 bottles of 200ml or total quantity of 1000ml) should be supplied with the equipment. j) Warranty includes the entire machine including probe. k) Standards, Safety & Training: a) Should be CE ("Conformite Europeene") from European Union notified body having 4 digit identification number or US FDA certification. Only product enlistment under US FDA will not be accepted. b) Warranty: 2 years; Extended Warranty: 3 year c) On site comprehensive training for field staff and support services till customer satisfaction with the system l) Documentation: a) User / Technical / Maintenance manuals to be supplied in English	73 No

B. K. S. by  
 Biomedical Engineer  
 District Health & Family Welfare Samity  
 Cooch Behar  
 17/02/2021

17/02/21  
 Dy. Chief Medical Officer of Health-1  
 Cooch Behar

17/02/21  
 Chief Medical Officer of Health  
 Cooch Behar

## Equipments specification for Suswastya Kendras (Non Cat Items)

4.	MEASURING TAPE ( MUAC)	<p>Measuring Tape (MUAC):</p> <p>a) Mid Upper Arm Circumference Tape - Child.</p> <p>b) MUAC tapes are predominately used to measure the upper arm circumference of children, helping identify malnutrition.</p> <p>c) Flexible &amp; Washable Material. Measuring in cm. Graduation - 1mm.</p> <p>d) Measuring Up To 30+ cm.</p> <p>e) Made from soft non bearable vinyl sheet.</p> <p>f) Colour Coded: 3-11.5cm (Red), 11.5-12.5cm (Yellow) 12.5 To 30+cm (Green).</p>	146 No
5.	MEASURING TAPE ( WC)	<p>Measuring Tape (WC):</p> <p>a) Application: Clinic</p> <p>b) Material: Plastic</p> <p>c) Packing Type: Box</p> <p>d) Width: 3 Metre</p> <p>e) Thickness: 9.5 mm</p>	146 No
6.	DENTAL PROB E WITH MIRROR SET	<p>Dental Probe with Mirror set:</p> <p>a) Materials: Made of High Quality Medical Grade Stainless Steel</p> <p>b) Set consists of: Mouth Mirror with handle, Probe Double Ended; Tweezers</p> <p>c) Mouth mirror: - Concave mirrors used by dentists because at a short range (object distances less than the focal length) they produce magnified, upright images. It is useful to have a magnified image of a tooth when you're looking for or repairing cavities, cracks, or other abnormalities</p> <p>d) Probe:- It is an instrument commonly used in the dental armamentarium. It is usually long, thin, and blunted at the end. The primary purpose of a periodontal probe is to measure pocket depths around a tooth in order to establish the state of health of the periodontium.</p> <p>e) Tweezers:- commonly used for retrieving objects from the mouth or placing objects into the mouth with somewhat similar functions with forceps.</p>	73 No
7.	EXAMINATION LAMP	<p>Examination Lamp:</p> <p>a) Light Source Type : LED</p> <p>b) Light Colour : White</p> <p>c) Usage/Application : Hospital</p> <p>d) Voltage : 200 – 250 V</p> <p>e) Wattage : 8 - 30 W</p> <p>f) Mounded : stand with an iron cast base</p> <p>g) Positioning : Mobile and adjust the size</p> <p>h) Arm : There is a flexible arm for easy movement of the lamp in all directions</p>	73 No
8.	TONGUE DEPRESSOR	<p>Tongue Depressor:</p> <p>a) Materials: Stainless Steel</p> <p>b) Surface Finish: Polished</p> <p>c) Reusable: Yes</p>	73 No
9.	MOUTH GAG	<p>Mouth Gag:</p> <p>a) Material: Stainless Steel</p> <p>b) Weight : upto 80 gm</p> <p>c) Usage/Application: Clinical</p> <p>d) Portable : Yes</p> <p>e) Rust proof</p> <p>f) Highly durable</p> <p>g) Smooth finish</p>	73 No
10.	SNELLEN VISION CHART	<p>Snellen Vision Chart:</p> <p>a) None reflecting, matte finish eye Charts, Designed for Vision testing at 20 Feet distance.</p> <p>b) Eye Chart measures 24 x 11 in.</p> <p>c) Plastic Washable Sheet.</p> <p>d) Wall Hanging Whole for easy hanging.</p>	73 No
11.	NEAR VISION CHART	<p>Near Vision Chart:</p> <p>a) Notations for vision testing at 40 cm (16in.)</p> <p>b) Visual Acuity range 8M to .20M (20/200, 6/120 to 20/10, 6/3 equiv.)</p> <p>c) 40 cm (16 in.) cord to maintain proper viewing distance</p> <p>d) Eye Chart Measures 18 X 23 cm (7 X 9 in.)</p>	73 No

*Diksha*  
 Biomedical Engineer  
 District Health & Family Welfare Samity  
 Cooch Behar  
 17/02/2021

*Dr. Chief Medical Officer of Health-1*  
 Cooch Behar

*17/02/21*  
 Chief Medical Officer of Health  
 Cooch Behar

## Equipments specification for Suswastya Kendras (Non Cat Items)

12.	COLOUR VISION CHART VI	<p>Colour Vision Chart:</p> <p>a) 38 Plate test Book.</p> <p>b) Accepted by leading authorities as a simple and accurate method for discovering congenital color blindness and red-green blindness, each in two forms, complete and incomplete.</p> <p>c) Includes four special plates for tests to determine the kind and degree of defect in color vision.</p> <p>d) Makes use of the peculiarity that in red-green blindness, blue and yellow appear remarkably bright compared with red and green.</p> <p>e) Color plates encased in specially designed album-type books for ease of handling. Detailed instructions included.</p>	73 No
13.	STDIOMETER	<p>Stadiometer:</p> <p>a) Height Stand Disassembles Into 4 Sections &amp; Easy To Transport In Disassembled Condition. Suitable For Height Measuring At Any Location</p> <p>b) Accurate, Portable, Durable And Light Weight With All Plastic Parts.</p> <p>c) Height Measuring Scale, Floor Model Without Weighing Scale.</p> <p>d) Measuring Range : 20 - 210Cm With 1Mm Graduation. Measuring In Inches And Centimeters</p>	73 No
14.	TUNING FORK	<p>Tuning Fork:</p> <p>a) A tuning fork is a metal instrument with a handle and two prongs or tines. Tuning forks, will vibrate at a set frequency to produce a musical tone when struck. The vibrations produced can be used to assess a person's ability to hear various sound frequencies</p> <p>b) A vibrating tuning fork held next to the ear or placed against the skull will stimulate the inner ear to vibrate, and can help determine if there is hearingloss</p> <p>c) A tuning fork is a metal instrument with a handle and two prongs or tines</p>	73 No

*S. K. S. for*  
17/06/2024  
**Biomedical Engineer**  
District Health & Family Welfare Samity  
Cooch Behar

*CS*  
17/6/24  
**Dy. Chief Medical Officer of Health-I**  
Cooch Behar

*CS*  
17/6/24  
**Chief Medical Officer of Health**  
Cooch Behar