



**Government of West Bengal
Office of the Medical Superintendent cum Vice Principal
Maharaja Jitendra Narayan Medical College & Hospital
Silver Jubilee Road, Cooch Behar**

No. -2726

Dated, Cooch Behar, the 17/11/2021

Quotation Notice

Sealed Quotations are invited from the bonafied Agency/Individual/Firm on behalf of the Medical Superintendent cum Vice Principal, MJN Medical College & Hospital, Cooch Behar for supply of below mentioned Item for ART Clinic of this hospital.

- [1] Table Top Centrifuge Machine - 01 No.
(Centrifuge Machine Specification Guideline attached)

The Quotation papers must be submitted in Tender Box from **17/11/2021**(except Sundays & Holidays) to **25/11/2021** between 10.00 AM to 3.00 PM to the office of the undersigned. The said quotation papers will be opened on **26/11/2021** at 11 AM at the office of the Medical Superintendent cum Vice Principal.

Terms & Conditions:-

1. The Quotation paper shall have to quote the offer rate of the item in figure and words clearly on an official pad of the supplier.
2. Rate should be quoted, clearly indicating all type of charges/GST etc. per unit as specified in above.
3. The supply order will be given to the successful bidder as selected by the selection committee and supply should be made within 15(Fifteen) days from the date of issue of the supply order.
4. The selected supplier will be fully responsible for supply & installation of the item to ART Clinic of MJN Government Medical College & Hospital or as directed to them.
5. It may also be noted that in case the quality of items to be supplied is not up to the mark in the opinion of the hospital authority, the supplier will be bound to take back the same without any claim and dispute etc. and the decision taken by the hospital authority shall be final and binding on the supplier.
6. The selection committee has the right to accept or reject any Quotation Paper without arising any reason.

**Medical Superintendent cum Vice Principal
MJN Medical College & Hospital
Dated, Cooch Behar, the 17/11/2021**

Memo No. – 2726/6(8)

Copy forwarded for information and necessary action to:-

1. The Sabhadhipati, Cooch Behar Zilla Parishad, Cooch Behar.
2. The District Magistrate, Cooch Behar.
3. The Principal, MJNMCH, Cooch Behar.
4. The Chairman, Cooch Behar Municipality.
5. The S.D.O., Cooch Behar Sadar.
6. The D.I.O., NIC, Cooch Behar.
7. Notice Board of this hospital.
8. Office file.

**Medical Superintendent cum Vice Principal
MJN Medical College & Hospital**

Specifications for Table Top Centrifuge Machine

Compact Table Top Laboratory Centrifuge with

1. Speed Regulator Maximum speed required at least 4000 rpm with increment of at least 100 rpm.
2. Safety lid lock.
3. Digital speed meter.
4. Timer for setting time between 0 to 60 with increment of at least 5 min.
5. Rotor angled head with capacity for at least 3-8ml blood tubes fit.
 - a) Adaptors each with capacity of 6 ml tube.
 - b) Adaptors each with capacity of 8 ml tube.
6. Facility for level adjustment and firm base to prevent walking and provide assure stability while equipment is in use.
7. Electrical compatibility
 - a) Must be compatible at 220V & 50/60Hz.
 - b) Power plug should fit on Indian system of electrical supply.
8. Should be able to accommodate minimum 12 tubes at one time.
9. The product must be CE/European/BIS approved.

Other Conditions:

1. The supplier will provide 5 years warranty that will include Comprehensive Annual Maintenance Contract (CAMC including all spare parts and repairs).
2. The manufacturer should be able to provide service of equipment across India within 24 hrs after receipt of breakdown report for the metro location and within 3 days for the non metro located instruments.
3. Purchase reserves the right to subject the equipments for independent evaluation of performance.
4. The company will provide installation qualification, operational qualification and performance qualification with log book for maintenance of the equipments at no extra cost.
5. Company will be responsible for Training of lab staff on operation of equipments.
6. Manual operation maintenance & part list with detailed specifications must be provided in original.



**Medical Superintendent cum Vice Principal
MJN Medical College & Hospital**