



COVID-19
URGENT

Government of West Bengal
Office of the Medical Superintendent cum Vice Principal
Maharaja Jitendra Narayan Medical College & Hospital
Silver Jubilee Road, Cooch Behar

No.: MJN/MCH/1515

Dated: Cooch Behar, the 25/06/2021

QUOTATION NOTICE

Sealed quotations are hereby invited on behalf of the Medical Superintendent cum Vice Principal, MJN Medical College & Hospital, Cooch Behar from reputed Firm/Manufacturer Company for supply of below mentioned **Eye Equipment for hospital patient services of COVID-19.**

- [1] **Direct Ophthalmoscope** - 02 No. (Specification of the Item stated in separate sheet)
- [2] **Liberman Eye Speculum** - 03 No.
- [3] **St. Martins Forceps** - 03 No.
- [4] **Eyeball Fixation Forceps** - 03 No.
- [5] **Barracker Needle Holder** - 03 No.
- [6] **Desmarre's Lid Retractor** - 04 No.
- [7] **Corneal Scissors'** - 03 No.
- [8] **Mule's Evisceration Scoop** - 02 No.
- [9] **Enucleation Scissor's (Straight-02 & Curved-02)** - 04 No.
- [10] **Squint Hook** - 03 No.

The quotation papers must be submitted in the "TENDER BOX" up to 30/06/2021 between 10.00 AM to 2.00 PM to the office of the undersigned. The said quotation papers will be opened on 01/07/2021 at 12:00 Noon at the office of the undersigned.

Terms & Conditions:-

1. The quotation shall have to quote the offer rate of the item in figure and words clearly on an official pad, addressing to the Medical Superintendent cum Vice Principal, M.J.N. Medical College & Hospital, Cooch Behar in their letter head pad.
2. Rate should be quoted, inclusive of all type of charges/GST etc. per unit as specified in above.
3. All the valid papers must be submitted along with quotation i.e. Copy of Pan Card/Trade Licence/GST IN certificate(Last Return)/Income Tax(Last Return).
4. The supply order will be given to the successful bidder as selected by the selection committee and supply should be made within 5(five) days from the date of issue of the supply order.
5. The selected party/Firm will be fully responsible for supply and installation of the items to Eye Department, MJN Medical College & Hospital.
6. It may also be noted that in case the quality of items to be supplied is not up to the mark in the opinion of the hospital authority, the supplier will be bound to take back the same without any claim and dispute etc. and the decision taken by the hospital authority shall be final and binding on the supplier.
7. The selection committee has the right to accept or reject any quotation without arising any reason.
8. Intending quotationers may please note that no change in the quotation, once submitted, shall be entertained and quotations found incomplete, incorrect, erased, overwritten, tampered will be summarily rejected.

Sd/-

Prof. Rajib Prasad

**Medical Superintendent cum Vice Principal
M. J. N. Medical College & Hospital**

Memo No.: MJN/MCH/1515/5(7)

Dated: Cooch Behar, the 25/06/2021

Copy forwarded for information and necessary action to:-

1. The Sabhadhipati, Cooch Behar Zilla Parishad.
2. The District Magistrate, Cooch Behar.
3. The Principal, MJNMCH, Cooch Behar with request to upload the notice in website.
4. The Accounts Officer, MJNMCH, Cooch Behar.
5. The D.I.O., NIC, Cooch Behar with request to upload the notice in www.coochbehar.nic.in
6. Notice Board of this hospital.
7. Office file.

**Medical Superintendent cum Vice Principal
M. J. N. Medical College & Hospital**

Technical Specification of Direct Ophthalmoscope

Exclusive dustproof design protects the precision optical components and ensures maintenance-free operation with an aluminium frame system. Optical components are flush-mounted on a cast aluminum frame, eliminating misalignment and making the instrument shock-resistant. Virtually unlimited hours of working life. Maintenance-free.

Aspherical optical system

Optical interface to precisely integrate the illumination with the optical system to ensure bright, homogeneous light during examinations.

Accommodates higher refractive errors with a 27 single lenses Rekoss Disc wheel avoiding extreme diopter jumps

+ in 1 D steps: +1 through + 10 15 20 40 D

- in 1 D steps: -1 through - 10 15 20 25 35 D

Stepless dimming from 3 % to 100 % with practical one-finger operation.

No stray light due to the recessed, multi-coated viewing window – eliminates distractions.

Precise one-hand control of the instrument. Ergonomic design offers flexibility while examining.

Suitable for large as well as small or undilated pupils.

Ergonomic design fits the orbita comfortably, and blocks out ambient light.

All metal battery handle to hold 2D cells

Isamb
4/9/20
HOD
Deptt. of Ophthalmology
Coochbehar Govt. Medical College & Hospital